

Facilities
Clinical Engineering Services
600 North Wolfe Street / Brady 101
Baltimore, MD 21287-2611
410-955-2100
410-502-6766 FAX



MEMORANDUM

To: Dr. Steven Yantis
Professor of Psychology
Johns Hopkins University

From: Roy Shipley, CBET
Associate Clinical Engineer
Clinical Engineering Services

Re: Your study entitled: "*EEG and fMRI Studies of Human Cognition*"
Protocol # NA_00005694
CES project # P-2022

Date: January 4, 2007

As requested by Joseph Gillen, I have tested the electrical safety of the MRA Inc. Video Eye Tracker Camera Interface. The device passed the AAMI/ANSI standard for external chassis leakage current. After reviewing the documentation for the study and inspecting the device, I am satisfied that the video eye-tracker presents no electrical or mechanical risk to study volunteers.

Clinical Engineering Services approves of the MRI compatible eye-tracker and the Compumedics Neuroscan model 8050 system unit (previously inspected and approved on October 19, 2006 for protocol # NA_00005004) for use in this study. If you have any questions, please contact me by phone at 5-2100 ext.106 or via E-mail royshi@jhmi.edu.

Cc: Jared Abrams
Daniel Caggiano
Andrew Currie
Joseph Gillen