

Facilities
Clinical Engineering Services
600 North Wolfe Street / Brady 101
Baltimore, MD 21287-2611
410-955-2100
410-502-6766 FAX



MEMORANDUM

To: Gwenn Smith, PhD
Professor
SOM Psychiatry and Behavioral Sciences

James J. Pekar, Ph.D.
Associate Professor of Radiology
Johns Hopkins University Manager and Research Coordinator,
F. M. Kirby Research Center for Functional Brain Imaging
Kennedy Krieger Institute

From: Roy Shipley, CBET
Associate Clinical Engineer
Clinical Engineering Services

Re: Protocol number: NA_00026190
Your study entitled: **“Serotonin Modulation in Mild Cognitive Impairment”**
CES project number: P-2641IRB

Date: October 13, 2010

As requested by Joseph Gillen, I have inspected the Philips 7.0 Tesla MR scanner that was recently installed at the Kennedy Krieger Institute’s Clinical Research Center. After seeing the installation, discussing the system specifications with Joe Gillen, and reviewing the FDA letter, I agree that the scanner represents a non-significant risk to study volunteers.

Clinical Engineering Services approves of the Philips 7T MR scanner for use in this study protocol. If you have any questions, please contact me by phone at 5-069 or by E-mail rshiple1@jhmi.edu.

Cc: Andrew Currie
Joseph Gillen